



Sterility Assurance Products
World Class and Globally Trusted

Albert Browne Ltd
A subsidiary of STERIS Corporation

23 January 2016

To Whom it May Concern

All chemical indicators manufactured by Albert Browne have no lead or other heavy metals added to them during their manufacture.

When used as intended, these products do not release any substances known to be toxic in sufficient quantities to cause a health hazard, during or after the sterilization process for which it is designated.

CLEANING

REF	DESCRIPTION	Box Qty
2304	Washer Disinfectant Test Soil	10 Pots
2312	STF Load Check	25 Indicators
2315	STF Load Check	100 Indicators
2370	Ninhydrin Protein Detection Test	4 Tests
2369	Bulk Ninhydrin Protein Detection Test	25 Tests
3369	ResiTEST Residual Protein Detection Test Kit	25 Tests
3371	ResiTEST Residual Protein Pen	25 Tests
HX601	HexaLumen Test Kit Indicators	60 Indicators

THERMAL DISINFECTION

REF	DESCRIPTION	Box Qty
2470	Des Check 93°C/10mins	100 Indicators
2471	Des Check 90°C/5mins	100 Indicators

BOWIE DICK/STEAM PENETRATION

REF	DESCRIPTION	Box Qty
2310	TST Bowie Dick Test Pack for sterilizers working at 121°C - 124°C for 8 to 8.3 mins	20 Packs
2352	TST Bowie Dick Test Pack for sterilizers working at 134°C - 137°C for up to 3.5 mins	20 Packs
2385	Sensor Chemical Indicator Test Sheet for use at 121°C - 137°C in a standard Towel Pack	36 Sheets
3780	TST Helix working at 132°C - 137°C for up to 3.5 mins	1 Helix/250 Indicators
AB100	AB100 Porous & Hollow Steam Penetration Test working at 134°C - 137°C for up to 3.5 mins	1 Device/400 Indicators
AB101	AB101 Porous & Hollow Steam Penetration Test working at 134°C - 137°C for up to 3.5 mins	1 Device/400 Indicators

LOAD CHECK

REF	DESCRIPTION	Box Qty
3777	TST Load Check working at 134°C for 4 mins / 121°C for 12 mins	1 Load Check/400 Indicators
3778	TST Load Check working at 134°C for 5.3mins / 121°C for 15 mins	1 Load Check/400 Indicators
3779	TST Load Check working at 134°C for 7 mins / 121°C for 20 mins	1 Load Check/400 Indicators
3783	TST Load Check working at 134°C for 3.5 mins	1 Load Check/400 Indicators
2445	Formaldehyde Load check	1 Load check/200 Indicators



STERILIZATION TST TYPE 6 INDICATORS

REF	DESCRIPTION	Box Qty
2317	TST Control™ Emulating Indicators 134°C/3.5 mins 'Flash'	200 Indicators
2340	TST Control™ Emulating Indicators 134°C/3.5 mins	100 Indicators
3727	TST Control™ Emulating Indicators 134°C/3.5 mins - Self Adhesive	100 Indicators
2341	TST Control™ Emulating Indicators 134°C/4 mins & 121°C/12 mins	100 Indicators
2302	TST Control™ Emulating Indicators 134°C/5.3mins & 121°C/15 mins	200 Indicators
2342	TST Control™ Emulating Indicators 134°C/5.3mins & 121°C/15 mins	100 Indicators
3726	TST Control™ Emulating Indicators 134°C/5.3mins & 121°C/15 mins - Self Adhesive	100 Indicators
3725	TST Control™ Emulating Indicators 134°C/7mins & 121/20mins - Self Adhesive	200 Indicators
3760	TST Control™ Emulating Indicators 134°C/7mins & 121°C/20mins	200 Indicators
3702	TST Control™ Emulating Indicators 134°C/9mins	200 Indicators
3706	TST Control™ Emulating Indicators 134°C/18mins	200 Indicators
3103	134/3.5 min Type 6 Chemical Indicators for use with AB123 Spiral Load Check Device	200 Indicators
3104	134/4.0 min Type 6 Chemical Indicators for use with AB123 Spiral Load Check Device	200 Indicators
3105	134/5.3 min Type 6 Chemical Indicators for use with AB123 Spiral Load Check Device	200 Indicators
3107	134/7.0 min Type 6 Chemical Indicators for use with AB123 Spiral Load Check Device	200 Indicators

STERILIZATION STEAM, FORMALDEHYDE & EO INDICATORS

REF	DESCRIPTION	Box Qty
2401	Formaldehyde Control Indicators	100 Indicators
2402	Formaldehyde Form Process Indicator Spots	250 Indicators
2563	EO MVI Indicators	250 Indicators
2420	EO Integrator	100 Indicators
2421	EO Process Indicator Spots	1000 Indicators
2501	Vapourstrip VH202 Indicator	200 Indicators
2551	Steam MVI Indicators	240 Indicators
3100	Steam Type 5 Integrating Indicators	250 Indicators

STERILIZATION CONTROL TUBES

REF	DESCRIPTION	Box Qty
7301	Sterilization Control Tube Type 1 – Black Spot	100 Tubes
7302	Sterilization Control Tube Type 2 – Yellow Spot	100 Tubes
7303	Sterilization Control Tube Type 3 – Green Spot	100 Tubes
7304	Sterilization Control Tube Type 4 – Blue Spot	100 Tubes
7305	Sterilization Control Tube Type 5 – White Spot	100 Tubes

DOCUMENTATION & RECORD KEEPING

REF	DESCRIPTION	Box Qty
3789	Documentation System Label English	24 Rolls / 500 Labels

ADHESIVE TAPES

REF	DESCRIPTION	Box Qty
0162	Tape With Steam Indicator 18mm x 50m	48 Rolls
0163	Tape With Steam Indicator 24mm x 50m	36 Rolls
0167	Tape With Steam Indicator 18mm x 50m Blue	48 Rolls
0168	Tape With Steam Indicator 24mm x 50m Blue	36 Rolls
0151	Tape With EO Indicator 18mm x 50m	48 Rolls

Yours sincerely



Richard Bancroft
Science & Technical Director