

**Aesculap AG
Regulatory Affairs**

Postfach 40
78501 Tuttlingen
Germany

Aesculap Chifa sp. z o.o.
ul. Tysiąclecia 14
64-300 Nowy Tomyśl
POLAND

Contact: Thomas Marquard
Fon: +49 7461 95-2361
Fax: +49 7461 95-2969
Email: thomas.marquard@aesculap.de
Internet: <http://www.aesculap.de>
Date: March 26, 2018


DECLARATION


We herewith certify that the products

Article No .	Description
PL569T	LIGATING CLIPS M/L 12/BOX
PL572T	LIGATURE CLIP 12 MAG.= 144 PCS.
PL574T	CHALLENGER TI-P SM-LIGAT.CLIPS 12 CARTR.
PL579T	CHALLENGER TI-P ML-LIGAT.CLIPS 12 CARTR.

are designed, manufactured and tested according to the defined and documented specifications and procedures. These devices are produced from Titanium Grade 1 according to the ISO 5832/2^{x)}. Therefore the products are non ferromagnetic and MR-Conditional up to 3 Tesla.

AESCULAP AG

i. V.

Dr. Stefan Willemsen
Team Leader Regulatory Affairs

i. A.

Sandra Bury
Regulatory Affairs

^{x)} ISO 5832/2: "Implants for surgery – Metallic materials – Part 2: Unalloyed titanium"

Chairman of Supervisory Board:
Prof. Dr. h.c. Ludwig Georg Braun

Executive Board:
Dr. Joachim Schulz
(Chairman)
Dr. Jens von Lackum

Corporate Office: Tuttlingen
Register Court: Stuttgart HRB 726261
VAT reg. no. DE812160059

WEEE-Reg.-No. DE 65109852

Bank Account:
Deutsche Bank AG Tuttlingen
BLZ 653 700 75 Konto 21 22 000 00
IBAN DE44 6537 0075 0212 2000 00
SWIFT / BIC DEUTDE33
Baden-Württembergische Bank
BLZ 600 501 01 Konto 487 1905
IBAN DE31 6005 0101 0004 8719 05
SWIFT / BIC SOLADE33

Address:
Aesculap AG
Am Aesculap-Platz
78532 Tuttlingen
Germany