

Certificate of Analysis

Material : 23373.320

Dichloromethane HiPerSolv CHROMANORM for HPLC

Batch : 13Z2387

Expires end of 06/2016

CHARACTERISTICS	SPECIFICATIONS	MEASURED VALUES
Assay (GC)	Min. 99,80 %	99,92 %
Water	Max. 0,0500 %	0,0010 %
Non-volatile residue	Max. 0,001 %	Max. 0,001 %
Acidity	Max. 0,0005 meq/g	Max. 0,0005 meq/g
Alkalinity	Max. 0,0002 meq/g	Max. 0,0002 meq/g
Transmittance (260 nm)	Min. 98,0 %	100,0 %
Transmittance (245 nm)	Min. 80,0 %	96,3 %
Transmittance (240 nm)	Min. 50,0 %	86,8 %
Conforms to BDH 15245	Passes test	Passes test

We certify that this batch conforms to the specifications listed above.

BDL : Below detected limit.

Dr. Olaf C. Fehr, Chief Scientist - Europe

VWR International

Document printed on 04.07.2013

This document has been produced electronically and is valid without a signature.